

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092424\
 Data File : VN084136.D
 Acq On : 24 Sep 2024 15:55
 Operator : JC\MD
 Sample : P4116-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT191D2-20240917

Quant Time: Sep 25 01:35:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	161540	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	285389	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	242176	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	96083	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	125072	51.348	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.700%	
35) Dibromofluoromethane	8.165	113	91439	50.048	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.100%	
50) Toluene-d8	10.565	98	343354	54.362	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.720%	
62) 4-Bromofluorobenzene	12.847	95	115062	44.155	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 88.300%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.371	101	13284	7.180	ug/l	99
12) 1,1-Dichloroethene	4.347	96	2744	1.570	ug/l #	76
27) cis-1,2-Dichloroethene	7.483	96	1916	0.862	ug/l	86
30) Chloroform	7.965	83	1725	0.440	ug/l	97
38) Carbon Tetrachloride	8.359	117	1812	0.581	ug/l #	70
44) Trichloroethene	9.353	130	217584	113.519	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092424\
Data File : VN084136.D
Acq On : 24 Sep 2024 15:55
Operator : JC\MD
Sample : P4116-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT191D2-20240917

Quant Time: Sep 25 01:35:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

