

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092424\
 Data File : VN084137.D
 Acq On : 24 Sep 2024 16:19
 Operator : JC\MD
 Sample : P4116-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE125D1-20240917

Quant Time: Sep 25 01:36:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	150195	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	262510	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	224764	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	86841	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	118312	52.241	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	104.480%	
35) Dibromofluoromethane	8.171	113	88011	52.370	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	104.740%	
50) Toluene-d8	10.565	98	328013	56.459	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	112.920%	
62) 4-Bromofluorobenzene	12.847	95	108429	45.236	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	90.480%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.383	101	3496	2.032	ug/l #	82
12) 1,1-Dichloroethene	4.342	96	1009	0.621	ug/l #	57
27) cis-1,2-Dichloroethene	7.483	96	3911	1.892	ug/l	90
30) Chloroform	7.965	83	2000	0.549	ug/l	94
44) Trichloroethene	9.353	130	258405	146.566	ug/l	94
64) Tetrachloroethene	11.100	164	10731	7.179	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092424\
 Data File : VN084137.D
 Acq On : 24 Sep 2024 16:19
 Operator : JC\MD
 Sample : P4116-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE125D1-20240917

Quant Time: Sep 25 01:36:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

