

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092424\
 Data File : VN084139.D
 Acq On : 24 Sep 2024 17:07
 Operator : JC\MD
 Sample : P4116-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE125D3-20240917

Quant Time: Sep 25 01:37:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	147645	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	256724	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	222246	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	90278	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	116605	52.377	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.760%	
35) Dibromofluoromethane	8.165	113	85583	52.073	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.140%	
50) Toluene-d8	10.565	98	316974	55.788	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.580%	
62) 4-Bromofluorobenzene	12.847	95	107851	46.009	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 92.020%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.377	101	260068	153.800	ug/l	99
27) cis-1,2-Dichloroethene	7.488	96	2804	1.380	ug/l	85
44) Trichloroethene	9.353	130	125842	72.986	ug/l	95
64) Tetrachloroethene	11.106	164	17221	11.652	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092424\
Data File : VN084139.D
Acq On : 24 Sep 2024 17:07
Operator : JC\MD
Sample : P4116-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE125D3-20240917

Quant Time: Sep 25 01:37:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

