

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092425\
 Data File : VN087938.D
 Acq On : 24 Sep 2025 11:04
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Sep 25 03:08:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	184258	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	390513	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	375440	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	169941	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	169146	54.547	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.100%
35) Dibromofluoromethane	8.147	113	143911	50.756	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.520%
50) Toluene-d8	10.547	98	490463	49.190	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.380%
62) 4-Bromofluorobenzene	12.829	95	164238	46.082	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092425\
Data File : VN087938.D
Acq On : 24 Sep 2025 11:04
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Sep 25 03:08:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration

