

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092425\
 Data File : VN087942.D
 Acq On : 24 Sep 2025 12:28
 Operator : JC\MD
 Sample : Q3159-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-245

Quant Time: Sep 25 03:09:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	178118	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	398125	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	409216	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	182614	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	182802	60.983	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 121.960%	
35) Dibromofluoromethane	8.153	113	148945	51.527	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.060%	
50) Toluene-d8	10.547	98	527167	51.860	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.720%	
62) 4-Bromofluorobenzene	12.829	95	175653	48.343	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 96.680%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	30584	24.798	ug/l	93
18) Methyl Acetate	5.030	43	20702	6.707	ug/l	98
20) Methylene Chloride	5.259	84	11311	4.531	ug/l #	88
37) Ethyl Acetate	7.559	43	11776	2.580	ug/l	99
43) Isopropyl Acetate	8.671	43	13635	1.898	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092425\
Data File : VN087942.D
Acq On : 24 Sep 2025 12:28
Operator : JC\MD
Sample : Q3159-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-245

Quant Time: Sep 25 03:09:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration

