

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092518\
 Data File : VN051483.D
 Acq On : 25 Sep 2018 14:26
 Operator : MD\SY
 Sample : J5015-01RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPDES-1RE

Quant Time: Sep 26 02:33:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N092518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Sep 26 01:19:41 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	98037	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	487000	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	407314	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	202814	30.06	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.20%
60) 4-Bromofluorobenzene	12.40	95	144391	24.15	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	80.50%
63) Toluene-d8	10.09	98	571155	29.25	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	97.50%

Target Compounds					Ovalue
25) Chloroform	7.38	83	11162	1.146 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092518\
Data File : VN051483.D
Acq On : 25 Sep 2018 14:26
Operator : MD\SY
Sample : J5015-01RE
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPDES-1RE

Quant Time: Sep 26 02:33:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N092518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Sep 26 01:19:41 2018
Response via : Initial Calibration

