

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092519\
 Data File : VN058351.D
 Acq On : 25 Sep 2019 9:17
 Operator : JC/SP
 Sample : VN0925WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0925WBL01

Quant Time: Sep 26 01:42:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	471128	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	759405	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	649042	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	266089	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	344260	52.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.92%	
35) Dibromofluoromethane	7.57	113	241842	52.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.78%	
50) Toluene-d8	10.08	98	967657	53.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.60%	
62) 4-Bromofluorobenzene	12.41	95	308774	46.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.42%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092519\
Data File : VN058351.D
Acq On : 25 Sep 2019 9:17
Operator : JC/SP
Sample : VN0925WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0925WBL01

Quant Time: Sep 26 01:42:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

