

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092519\
 Data File : VN058356.D
 Acq On : 25 Sep 2019 11:07
 Operator : JC/SP
 Sample : PB123322TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB123322TB

Quant Time: Sep 26 05:55:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	427152	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	702306	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	600171	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	245931	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	313925	52.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.54%	
35) Dibromofluoromethane	7.57	113	218016	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
50) Toluene-d8	10.08	98	889614	53.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.96%	
62) 4-Bromofluorobenzene	12.41	95	287041	46.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.90%	
Target Compounds						
16) Acetone	3.80	43	15382	7.617	ug/l	97
20) Methylene Chloride	4.53	84	23800	6.539	ug/l	95
43) Isopropyl Acetate	8.15	43	97509	9.933	ug/l #	92
95) Naphthalene	15.14	128	19830	1.688	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092519\
Data File : VN058356.D
Acq On : 25 Sep 2019 11:07
Operator : JC/SP
Sample : PB123322TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB123322TB

Quant Time: Sep 26 05:55:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

