

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092520\
 Data File : VN063679.D
 Acq On : 25 Sep 2020 15:31
 Operator : JC/MD
 Sample : L4119-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE126D1-20200921

Quant Time: Sep 28 08:11:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	206713	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	388636	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	395558	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	161664	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	179028	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
35) Dibromofluoromethane	7.55	113	130631	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
50) Toluene-d8	10.06	98	503618	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
62) 4-Bromofluorobenzene	12.38	95	193951	52.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.60%	
Target Compounds						
3) Chloromethane	2.04	50	1068	0.345	ug/l #	82
44) Trichloroethene	8.80	130	105387	36.762	ug/l	98
64) Tetrachloroethene	10.60	164	3012	1.118	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092520\
Data File : VN063679.D
Acq On : 25 Sep 2020 15:31
Operator : JC/MD
Sample : L4119-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE126D1-20200921

Quant Time: Sep 28 08:11:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

