

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
 Data File : VN068661.D
 Acq On : 25 Sep 2021 11:47
 Operator : JC/MD
 Sample : VN0925WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0925WBL01

Quant Time: Sep 27 04:41:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	503449	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	818446	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	906014	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	266545	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	384822	48.959	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.920%
35) Dibromofluoromethane	8.024	113	290605	53.095	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.180%
50) Toluene-d8	10.443	98	1274403	54.903	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.800%
62) 4-Bromofluorobenzene	12.733	95	364840	45.958	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
Data File : VN068661.D
Acq On : 25 Sep 2021 11:47
Operator : JC/MD
Sample : VN0925WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0925WBL01

Quant Time: Sep 27 04:41:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

