

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
 Data File : VN068665.D
 Acq On : 25 Sep 2021 14:13
 Operator : JC/MD
 Sample : PB139230ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB139230ZHE#01

Manual Integrations
 APPROVED

MMDadoda
 9/27/2021 11:00:44 AM

Quant Time: Sep 27 04:43:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	562534	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	919246	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	1077970	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	383335	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	395936	44.984	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	89.960%		
35) Dibromofluoromethane	8.021	113	299546	48.654	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.300%		
50) Toluene-d8	10.443	98	1161542	44.554	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	89.100%		
62) 4-Bromofluorobenzene	12.734	95	356376	39.970	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	79.940%		
Target Compounds						
16) Acetone	4.291	43	70109	11.045	ug/l	Qvalue 99
20) Methylene Chloride	5.103	84	25683m	1.949	ug/l	
43) Isopropyl Acetate	8.563	43	53128	2.947	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
Data File : VN068665.D
Acq On : 25 Sep 2021 14:13
Operator : JC/MD
Sample : PB139230ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB139230ZHE#01

Manual Integrations
APPROVED

MMDadoda
9/27/2021 11:00:44 AM

Quant Time: Sep 27 04:43:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

