

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
 Data File : VN068666.D
 Acq On : 25 Sep 2021 14:38
 Operator : JC/MD
 Sample : PB139230ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB139230ZHE#02

Quant Time: Sep 27 04:43:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	638610	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	861189	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	794017	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	406973	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	388216	38.684	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	77.360%
35) Dibromofluoromethane	8.024	113	357618	62.246	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	124.500%
50) Toluene-d8	10.443	98	1404799	57.517	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	115.040%
62) 4-Bromofluorobenzene	12.733	95	509845	61.037	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	122.080%
Target Compounds						
					Qvalue	
16) Acetone	4.288	43	66326	7.201	ug/l	97
20) Methylene Chloride	5.103	84	25051	1.383	ug/l	92
43) Isopropyl Acetate	8.563	43	58324	3.453	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
 Data File : VN068666.D
 Acq On : 25 Sep 2021 14:38
 Operator : JC/MD
 Sample : PB139230ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB139230ZHE#02

Quant Time: Sep 27 04:43:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

