

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
 Data File : VN068667.D
 Acq On : 25 Sep 2021 15:03
 Operator : JC/MD
 Sample : PB139230ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB139230ZHE#03

Manual Integrations
 APPROVED

MMDadoda
 9/27/2021 11:00:47 AM

Quant Time: Sep 27 04:43:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	503317	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	835245	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	765892	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	373053	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	372308	47.340	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.680%		
35) Dibromofluoromethane	8.024	113	278414	49.790	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.580%		
50) Toluene-d8	10.446	98	1086051	45.848	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	91.700%		
62) 4-Bromofluorobenzene	12.733	95	335196	41.375	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	82.740%		
Target Compounds					Qvalue	
16) Acetone	4.291	43	105713	26.854	ug/l	96
20) Methylene Chloride	5.109	84	24488m	2.213	ug/l	
25) 2-Butanone	7.343	43	24456	5.105	ug/l	100
43) Isopropyl Acetate	8.563	43	59283	3.619	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
Data File : VN068667.D
Acq On : 25 Sep 2021 15:03
Operator : JC/MD
Sample : PB139230ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB139230ZHE#03

Manual Integrations
APPROVED

MMDadoda
9/27/2021 11:00:47 AM

Quant Time: Sep 27 04:43:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

