

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
 Data File : VN068669.D
 Acq On : 25 Sep 2021 15:53
 Operator : JC/MD
 Sample : PB139323TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB139323TB

Quant Time: Sep 27 04:44:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	515459	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	838255	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	838306	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	363104	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	397189	49.366	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.740%
35) Dibromofluoromethane	8.024	113	296459	52.881	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.760%
50) Toluene-d8	10.443	98	1277644	53.742	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.480%
62) 4-Bromofluorobenzene	12.734	95	578483	71.149	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	142.300%#
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	109770	27.395	ug/l	97
20) Methylene Chloride	5.101	84	27315	2.595	ug/l	90
25) 2-Butanone	7.337	43	26746	5.452	ug/l	100
43) Isopropyl Acetate	8.566	43	62181	3.782	ug/l	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
Data File : VN068669.D
Acq On : 25 Sep 2021 15:53
Operator : JC/MD
Sample : PB139323TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB139323TB

Quant Time: Sep 27 04:44:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

