

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
 Data File : VN068673.D
 Acq On : 25 Sep 2021 17:35
 Operator : JC/MD
 Sample : PB139363ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB139363ZHE#01

Quant Time: Sep 27 04:46:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	508637	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	877307	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	764249	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	272234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	336552	42.215	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	84.420%		
35) Dibromofluoromethane	8.024	113	253018	42.958	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	85.920%		
50) Toluene-d8	10.443	98	1033422	41.534	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	83.060%		
62) 4-Bromofluorobenzene	12.734	95	305452	35.896	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	71.800%		
Target Compounds						
16) Acetone	4.291	43	107292	27.022	ug/l	Qvalue 94
20) Methylene Chloride	5.101	84	30427	3.196	ug/l	98
25) 2-Butanone	7.343	43	25150	5.195	ug/l	100
43) Isopropyl Acetate	8.563	43	66532	3.866	ug/l	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
Data File : VN068673.D
Acq On : 25 Sep 2021 17:35
Operator : JC/MD
Sample : PB139363ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB139363ZHE#01

Quant Time: Sep 27 04:46:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

