

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
 Data File : VN068678.D
 Acq On : 25 Sep 2021 19:41
 Operator : JC/MD
 Sample : M3935-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Sep 27 04:47:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	540202	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	865949	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	916381	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	332774	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	373549	44.173	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	88.340%
35) Dibromofluoromethane	8.024	113	273102	47.060	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.120%
50) Toluene-d8	10.443	98	1282746	52.231	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.460%
62) 4-Bromofluorobenzene	12.733	95	395208	47.053	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.100%
Target Compounds						
						Qvalue
16) Acetone	4.291	43	112334	26.468	ug/l	98
20) Methylene Chloride	5.106	84	34709	3.586	ug/l	94
25) 2-Butanone	7.340	43	26464	5.147	ug/l	98
43) Isopropyl Acetate	8.563	43	62328	3.670	ug/l	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
Data File : VN068678.D
Acq On : 25 Sep 2021 19:41
Operator : JC/MD
Sample : M3935-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Sep 27 04:47:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

