

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
 Data File : VN068680.D
 Acq On : 25 Sep 2021 20:31
 Operator : JC/MD
 Sample : PB139290TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB139290TB

Quant Time: Sep 27 04:48:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	579910	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	960544	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	1102086	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	476238	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	451421	49.883	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.760%
35) Dibromofluoromethane	8.024	113	335781	52.259	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.520%
50) Toluene-d8	10.448	98	1475356	54.158	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.320%
62) 4-Bromofluorobenzene	12.733	95	417859	44.850	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.700%
Target Compounds						
						Qvalue
16) Acetone	4.291	43	108357	22.563	ug/l	93
20) Methylene Chloride	5.106	84	31192	2.665	ug/l	94
43) Isopropyl Acetate	8.563	43	64313	3.413	ug/l	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
Data File : VN068680.D
Acq On : 25 Sep 2021 20:31
Operator : JC/MD
Sample : PB139290TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB139290TB

Quant Time: Sep 27 04:48:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

