

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
 Data File : VN068694.D
 Acq On : 26 Sep 2021 4:29
 Operator : JC/MD
 Sample : M3895-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 210922013-01-TRIP-BLANK

Quant Time: Sep 27 07:24:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N092521W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Sep 27 07:10:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.662	128	69826	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.971	114	388802	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.749	117	334326	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	177386	30.900	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.000%
60) 4-Bromofluorobenzene	12.733	95	127878	25.039	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	83.467%
63) Toluene-d8	10.443	98	502675	33.974	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	113.233%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092521\
Data File : VN068694.D
Acq On : 26 Sep 2021 4:29
Operator : JC/MD
Sample : M3895-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
210922013-01-TRIP-BLANK

Quant Time: Sep 27 07:24:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N092521W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Sep 27 07:10:18 2021
Response via : Initial Calibration

