

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092523\
 Data File : VN079338.D
 Acq On : 25 Sep 2023 11:49
 Operator : JC\MD
 Sample : VN0925WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0925WBL01

Quant Time: Sep 26 06:18:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	160815	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	296965	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	258687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	73146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	133178	61.348	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 122.700%	
35) Dibromofluoromethane	8.171	113	102911	52.098	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.200%	
50) Toluene-d8	10.571	98	338681	46.694	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 93.380%	
62) 4-Bromofluorobenzene	12.853	95	107315	46.384	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.760%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092523\
Data File : VN079338.D
Acq On : 25 Sep 2023 11:49
Operator : JC\MD
Sample : VN0925WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0925WBL01

Quant Time: Sep 26 06:18:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

