

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
 Data File : VN084150.D
 Acq On : 25 Sep 2024 10:47
 Operator : JC\MD
 Sample : VN0925MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0925MBL01

Quant Time: Sep 26 01:31:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	158728	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	274097	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	231707	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	93531	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	120707	50.433	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.860%	
35) Dibromofluoromethane	8.165	113	90725	51.703	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.400%	
50) Toluene-d8	10.565	98	329902	54.384	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.760%	
62) 4-Bromofluorobenzene	12.847	95	113423	45.319	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 90.640%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
Data File : VN084150.D
Acq On : 25 Sep 2024 10:47
Operator : JC\MD
Sample : VN0925MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0925MBL01

Quant Time: Sep 26 01:31:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

