

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
 Data File : VN084151.D
 Acq On : 25 Sep 2024 11:11
 Operator : JC\MD
 Sample : VN0925WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0925WBL01

Quant Time: Sep 26 01:31:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	169687	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	292354	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	243491	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	100585	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	127336	49.767	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.540%
35) Dibromofluoromethane	8.165	113	95293	50.915	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.820%
50) Toluene-d8	10.565	98	348398	53.846	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.700%
62) 4-Bromofluorobenzene	12.847	95	120193	45.025	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
Data File : VN084151.D
Acq On : 25 Sep 2024 11:11
Operator : JC\MD
Sample : VN0925WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0925WBL01

Quant Time: Sep 26 01:31:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

