

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
 Data File : VN084155.D
 Acq On : 25 Sep 2024 13:23
 Operator : JC\MD
 Sample : P4116-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT15811-20240916

Quant Time: Sep 26 01:33:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	166393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	286124	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	247589	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	104875	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	124291	49.539	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.080%
35) Dibromofluoromethane	8.171	113	92868	50.700	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.400%
50) Toluene-d8	10.565	98	349747	55.231	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.460%
62) 4-Bromofluorobenzene	12.847	95	122060	46.720	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
Data File : VN084155.D
Acq On : 25 Sep 2024 13:23
Operator : JC\MD
Sample : P4116-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT15811-20240916

Quant Time: Sep 26 01:33:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

