

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
 Data File : VN084164.D
 Acq On : 25 Sep 2024 16:59
 Operator : JC\MD
 Sample : P4116-18DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE131D3-20240917DL

Quant Time: Sep 26 01:35:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	170640	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	287714	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	256290	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	108779	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	126373	49.115	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.220%
35) Dibromofluoromethane	8.165	113	95707	51.961	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.920%
50) Toluene-d8	10.565	98	361266	56.735	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	113.480%#
62) 4-Bromofluorobenzene	12.847	95	126491	48.148	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.300%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.371	101	78439	40.136	ug/l	99
44) Trichloroethene	9.353	130	8365	4.329	ug/l	97
64) Tetrachloroethene	11.100	164	6179	3.625	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
Data File : VN084164.D
Acq On : 25 Sep 2024 16:59
Operator : JC\MD
Sample : P4116-18DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE131D3-20240917DL

Quant Time: Sep 26 01:35:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

