

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
 Data File : VN084165.D
 Acq On : 25 Sep 2024 17:23
 Operator : JC\MD
 Sample : P4116-08DL 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE122D2-20240916DL

Quant Time: Sep 26 01:36:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	161998	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	290894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	253037	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	109699	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	125416	51.343	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.680%
35) Dibromofluoromethane	8.165	113	94984	51.005	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.000%
50) Toluene-d8	10.565	98	354514	55.066	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.140%
62) 4-Bromofluorobenzene	12.847	95	125706	47.327	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.660%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.353	130	66356	33.964	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
Data File : VN084165.D
Acq On : 25 Sep 2024 17:23
Operator : JC\MD
Sample : P4116-08DL 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE122D2-20240916DL

Quant Time: Sep 26 01:36:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

