

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
 Data File : VN084166.D
 Acq On : 25 Sep 2024 17:47
 Operator : JC\MD
 Sample : P4116-12DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT191D2-20240917DL

Quant Time: Sep 26 01:36:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	155642	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	280711	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	246655	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	105815	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	121624	51.824	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.640%			
35) Dibromofluoromethane	8.165	113	90694	50.468	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.940%			
50) Toluene-d8	10.565	98	338339	54.460	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 108.920%			
62) 4-Bromofluorobenzene	12.847	95	122062	47.622	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 95.240%			
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.377	101	2540	1.425	ug/l	94
44) Trichloroethene	9.353	130	42664	22.630	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
Data File : VN084166.D
Acq On : 25 Sep 2024 17:47
Operator : JC\MD
Sample : P4116-12DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT191D2-20240917DL

Quant Time: Sep 26 01:36:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

