

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
 Data File : VN084167.D
 Acq On : 25 Sep 2024 18:11
 Operator : JC\MD
 Sample : P4116-16DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE131D1-20240917DL

Quant Time: Sep 26 01:36:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153846	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	273267	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	245973	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	111803	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	121922	52.558	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.120%	
35) Dibromofluoromethane	8.171	113	91560	52.337	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.680%	
50) Toluene-d8	10.565	98	337320	55.775	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.560%	
62) 4-Bromofluorobenzene	12.847	95	124432	49.869	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 99.740%	
Target Compounds						
						Qvalue
44) Trichloroethene	9.353	130	73517	40.057	ug/l	97
64) Tetrachloroethene	11.106	164	4623	2.826	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
Data File : VN084167.D
Acq On : 25 Sep 2024 18:11
Operator : JC\MD
Sample : P4116-16DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE131D1-20240917DL

Quant Time: Sep 26 01:36:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

