

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
 Data File : VN084168.D
 Acq On : 25 Sep 2024 18:35
 Operator : JC\MD
 Sample : P4116-22
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE132D4-20240918

Quant Time: Sep 26 01:36:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	155260	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	274137	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	247955	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	104708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	118336	50.547	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.100%	
35) Dibromofluoromethane	8.165	113	91600	52.194	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.380%	
50) Toluene-d8	10.565	98	337540	55.635	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.260%	
62) 4-Bromofluorobenzene	12.847	95	120274	48.049	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 96.100%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.365	101	6928	3.896	ug/l	# 30
44) Trichloroethene	9.353	130	282106	153.223	ug/l	93
64) Tetrachloroethene	11.100	164	6393	3.877	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092524\
Data File : VN084168.D
Acq On : 25 Sep 2024 18:35
Operator : JC\MD
Sample : P4116-22
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE132D4-20240918

Quant Time: Sep 26 01:36:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

