

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092525\
 Data File : VN087949.D
 Acq On : 25 Sep 2025 12:37
 Operator : JC\MD
 Sample : VN0925MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0925MBL01

Quant Time: Sep 26 07:29:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	174365	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	385843	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	371436	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	167826	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	168132	57.296	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.600%
35) Dibromofluoromethane	8.147	113	143885	51.361	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.720%
50) Toluene-d8	10.547	98	486851	49.419	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.840%
62) 4-Bromofluorobenzene	12.829	95	155192	44.071	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092525\
Data File : VN087949.D
Acq On : 25 Sep 2025 12:37
Operator : JC\MD
Sample : VN0925MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0925MBL01

Quant Time: Sep 26 07:29:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration

