

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092525\
 Data File : VN087955.D
 Acq On : 25 Sep 2025 15:24
 Operator : JC\MD
 Sample : Q3177-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EME-TS14-02

Quant Time: Sep 26 07:32:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	173667	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	381223	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	364834	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	168811	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	160415	54.886	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.780%	
35) Dibromofluoromethane	8.147	113	138977	50.210	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.420%	
50) Toluene-d8	10.547	98	461273	47.390	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 94.780%	
62) 4-Bromofluorobenzene	12.829	95	153204	44.034	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 88.060%	
Target Compounds						
					Qvalue	
16) Acetone	4.418	43	9249	7.691	ug/l #	87
20) Methylene Chloride	5.259	84	22431	9.216	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092525\
Data File : VN087955.D
Acq On : 25 Sep 2025 15:24
Operator : JC\MD
Sample : Q3177-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EME-TS14-02

Quant Time: Sep 26 07:32:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration

