

Data File : VN051500.D
Acq On : 25 Sep 2018 22:26
Operator : MD\SY
Sample : J5065-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-111

Manual Integrations
APPROVED

mmdadoda
9/30/2018 9:54:02 AM

Quant Time: Oct 05 04:07:25 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	644712	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	985949	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	837402	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	354018	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	439020	54.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.72%	
35) Dibromofluoromethane	7.59	113	389773	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
50) Toluene-d8	10.09	98	1401160	47.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.38%	
62) 4-Bromofluorobenzene	12.40	95	405414	42.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.42%	
Target Compounds						
16) Acetone	3.82	43	95710	32.387	ug/l	99
27) cis-1,2-Dichloroethene	6.83	96	5049	0.627	ug/l	92
30) Chloroform	7.38	83	21489	1.562	ug/l	95
37) Ethyl Acetate	6.94	43	11353m	1.729	ug/l	
60) Dibromochloromethane	10.90	129	5122	0.590	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN051500.D
Acq On : 25 Sep 2018 22:26
Operator : MD\SY
Sample : J5065-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-111

Manual Integrations
APPROVED

mmdadoda
9/30/2018 9:54:02 AM

Quant Time: Oct 05 04:07:25 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

