

Data File : VN051503.D
Acq On : 25 Sep 2018 23:40
Operator : MD\SY
Sample : J5065-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-124

Manual Integrations
APPROVED

apatel
10/1/2018 12:20:14 PM

Quant Time: Oct 05 04:08:26 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	653761	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	994499	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	809463	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	327774	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	438339	53.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.04%	
35) Dibromofluoromethane	7.59	113	382889	48.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.30%	
50) Toluene-d8	10.09	98	1393178	46.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.04%	
62) 4-Bromofluorobenzene	12.40	95	385457	39.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.58%	
Target Compounds						
25) 2-Butanone	6.84	43	18155	6.477	ug/l	91
27) cis-1,2-Dichloroethene	6.84	96	24238m	2.969	ug/l	
30) Chloroform	7.37	83	25515	1.829	ug/l	98
37) Ethyl Acetate	6.94	43	12971	1.959	ug/l #	90
44) Trichloroethene	8.84	130	3872	0.431	ug/l	92
52) Toluene	10.16	92	33222	1.581	ug/l	100
64) Tetrachloroethene	10.63	164	8770	1.194	ug/l	92
67) Ethyl Benzene	11.52	91	12770	0.365	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN051503.D
Acq On : 25 Sep 2018 23:40
Operator : MD\SY
Sample : J5065-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-124

Manual Integrations
APPROVED

apatel
10/1/2018 12:20:14 PM

Quant Time: Oct 05 04:08:26 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

