

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092618\
 Data File : VN051524.D
 Acq On : 26 Sep 2018 9:06
 Operator : MD\SY
 Sample : PB113167ZHE#13
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113167ZHE#13

Quant Time: Sep 26 11:33:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	611830	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	990914	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	822830	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	282158	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	434689	56.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.44%	
35) Dibromofluoromethane	7.59	113	378031	47.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.42%	
50) Toluene-d8	10.09	98	1359348	45.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.12%	
62) 4-Bromofluorobenzene	12.40	95	365032	37.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.64%	
Target Compounds						
18) Methyl Acetate	4.33	43	14678	2.515	ug/l #	90
20) Methylene Chloride	4.55	84	8743	1.153	ug/l	94
43) Isopropyl Acetate	8.17	43	341264	28.605	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092618\
Data File : VN051524.D
Acq On : 26 Sep 2018 9:06
Operator : MD\SY
Sample : PB113167ZHE#13
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB113167ZHE#13

Quant Time: Sep 26 11:33:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

