

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092618\
 Data File : VN051525.D
 Acq On : 26 Sep 2018 9:31
 Operator : MD\SY
 Sample : PB113167ZHE#14
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113167ZHE#14

Quant Time: Sep 26 11:46:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	598965	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	969231	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	812241	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	282141	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	430137	57.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.66%	
35) Dibromofluoromethane	7.59	113	373567	48.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.40%	
50) Toluene-d8	10.09	98	1343452	46.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.06%	
62) 4-Bromofluorobenzene	12.40	95	351693	37.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.50%	
Target Compounds						
18) Methyl Acetate	4.33	43	16028	2.805	ug/l #	74
20) Methylene Chloride	4.55	84	8226	1.108	ug/l	97
43) Isopropyl Acetate	8.17	43	335265	28.730	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092618\
Data File : VN051525.D
Acq On : 26 Sep 2018 9:31
Operator : MD\SY
Sample : PB113167ZHE#14
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB113167ZHE#14

Quant Time: Sep 26 11:46:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

