

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092618\
 Data File : VN051526.D
 Acq On : 26 Sep 2018 9:56
 Operator : MD\SY
 Sample : PB113167ZHE#15
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113167ZHE#15

Quant Time: Sep 26 11:52:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	610303	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1001292	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	850413	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	311058	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	448802	58.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.40%	
35) Dibromofluoromethane	7.59	113	384230	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
50) Toluene-d8	10.09	98	1388616	46.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.12%	
62) 4-Bromofluorobenzene	12.40	95	378246	38.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.56%	
Target Compounds						
18) Methyl Acetate	4.34	43	15059	2.586	ug/l #	80
20) Methylene Chloride	4.55	84	4838	0.640	ug/l	89
43) Isopropyl Acetate	8.17	43	348996	28.950	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092618\
Data File : VN051526.D
Acq On : 26 Sep 2018 9:56
Operator : MD\SY
Sample : PB113167ZHE#15
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB113167ZHE#15

Quant Time: Sep 26 11:52:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

