

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092619\
 Data File : VN058375.D
 Acq On : 26 Sep 2019 11:17
 Operator : JC/SP
 Sample : K5042-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 55-ST-21

Quant Time: Sep 27 11:23:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N092419W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Sep 25 01:36:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.18	128	76163	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.57	114	435128	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	374385	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.01	65	208494	30.97	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	103.23%
60) 4-Bromofluorobenzene	12.41	95	172332	26.60	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	88.67%
63) Toluene-d8	10.08	98	543881	30.44	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.47%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092619\
Data File : VN058375.D
Acq On : 26 Sep 2019 11:17
Operator : JC/SP
Sample : K5042-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
55-ST-21

Quant Time: Sep 27 11:23:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N092419W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Sep 25 01:36:25 2019
Response via : Initial Calibration

