

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092619\
 Data File : VN058378.D
 Acq On : 26 Sep 2019 12:23
 Operator : JC/SP
 Sample : PB123389TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB123389TB

Quant Time: Sep 27 04:07:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	470739	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	776806	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	678947	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	270799	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	350632	53.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.96%	
35) Dibromofluoromethane	7.58	113	243959	51.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.32%	
50) Toluene-d8	10.08	98	990153	53.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.62%	
62) 4-Bromofluorobenzene	12.41	95	323109	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
Target Compounds						
16) Acetone	3.81	43	12103	5.438	ug/l	93
20) Methylene Chloride	4.53	84	38945	9.817	ug/l	95
43) Isopropyl Acetate	8.15	43	113877	10.488	ug/l #	90
80) 1,3,5-Trimethylbenzene	12.74	105	14680	1.024	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092619\
Data File : VN058378.D
Acq On : 26 Sep 2019 12:23
Operator : JC/SP
Sample : PB123389TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB123389TB

Quant Time: Sep 27 04:07:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

