

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092619\
 Data File : VN058370.D
 Acq On : 26 Sep 2019 9:06
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Sep 27 03:33:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N092419W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Sep 25 01:36:25 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	93	0.00
2 M	Dichlorodifluoromethane	1.804	1.935	-7.3	97	0.00
3 M	Chloromethane	2.083	2.130	-2.3	96	0.00
4 M	Vinyl Chloride	2.079	2.122	-2.1	94	0.00
5 M	Bromomethane	1.129	1.168	-3.5	103	0.00
6 M	Chloroethane	1.228	1.267	-3.2	94	0.00
7 M	Trichlorofluoromethane	2.447	2.557	-4.5	94	0.00
8 T	Diethyl Ether	1.088	1.105	-1.6	96	0.00
9	1,1,2-Trichlorotrifluoroeth	1.685	1.738	-3.1	96	0.00
10 M	1,1-Dichloroethene	1.647	1.671	-1.5	94	0.00
11	Methyl Iodide	1.689	1.806	-6.9	112	0.00
12	Methyl Acetate	2.193	2.075	5.4	92	0.00
13 M	Acrolein	0.062	0.059	4.8	86	0.00
14 M	Acrylonitrile	0.906	0.883	2.5	93	0.00
15 M	Acetone	0.251	0.356	-41.8#	125	0.00
16 M	Carbon Disulfide	4.577	4.392	4.0	96	0.00
17	Allyl chloride	3.076	3.043	1.1	94	0.00
18 M	Methylene Chloride	1.943	1.999	-2.9	96	0.00
19 M	trans-1,2-Dichloroethene	1.846	1.862	-0.9	95	0.00
20 T	Diisopropyl ether	6.555	6.522	0.5	95	0.00
21 M	1,1-Dichloroethane	3.668	3.752	-2.3	97	0.00
22 M	cis-1,2-Dichloroethene	2.145	2.164	-0.9	95	0.00
23 M	tert-Butyl Alcohol	0.334	0.301	9.9	87	0.00
24 M	Methyl tert-Butyl Ether	5.912	5.759	2.6	92	0.00
25 M	Chloroform	3.714	3.751	-1.0	96	0.00
26	Cyclohexane	3.428	3.380	1.4	94	0.00
27 s	1,2-Dichloroethane-d4	2.652	2.689	-1.4	93	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
29	1,1-Dichloropropene	0.490	0.486	0.8	94	0.00
30 M	2-Butanone	0.222	0.245	-10.4	104	0.00
31	2,2-Dichloropropane	0.504	0.519	-3.0	98	0.00
32 M	1,1,1-Trichloroethane	0.528	0.535	-1.3	96	0.00
33 M	Carbon Tetrachloride	0.449	0.447	0.4	95	0.00
34 M	Benzene	1.389	1.404	-1.1	94	0.00
35	Methacrylonitrile	0.191	0.167	12.6	85	-0.01
36 M	1,2-Dichloroethane	0.544	0.533	2.0	93	0.00
37 M	Trichloroethene	0.348	0.347	0.3	94	0.00
38	Methylcyclohexane	0.592	0.581	1.9	95	0.00
39 M	1,2-Dichloropropane	0.379	0.382	-0.8	96	0.00
40	Dibromomethane	0.239	0.240	-0.4	95	0.00
41 M	Bromodichloromethane	0.473	0.472	0.2	97	0.00
42 M	Vinyl Acetate	0.849	0.909	-7.1	94	0.00
43	Ethyl Acetate	0.441	0.446	-1.1	97	0.00
44	Isopropyl Acetate	0.790	0.728	7.8	88	0.00
45 T	1,4-Dioxane	0.007	0.006#	14.3	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092619\
 Data File : VN058370.D
 Acq On : 26 Sep 2019 9:06
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Sep 27 03:33:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N092419W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Sep 25 01:36:25 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.375	0.348	7.2	91	0.00
47	n-amyl Acetate	0.688	0.623	9.4	90	0.00
48 M	t-1,3-Dichloropropene	0.517	0.494	4.4	96	0.00
49 T	cis-1,3-Dichloropropene	0.575	0.564	1.9	95	0.00
50 M	1,1,2-Trichloroethane	0.334	0.332	0.6	94	0.00
51	Ethyl methacrylate	0.529	0.491	7.2	91	0.00
52	1,3-Dichloropropane	0.594	0.580	2.4	92	0.00
53 M	Dibromochloromethane	0.336	0.318	5.4	95	0.00
54 M	1,2-Dibromoethane	0.341	0.332	2.6	94	0.00
55 M	2-Chloroethyl vinyl ether	0.224	0.201	10.3	86	0.00
56 M	Bromoform	0.212	0.190	10.4	95	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
58 M	4-Methyl-2-Pentanone	0.470	0.487	-3.6	90	0.00
59 M	2-Hexanone	0.355	0.383	-7.9	97	0.00
60 S	4-Bromofluorobenzene	0.519	0.505	2.7	92	0.00
61 M	Tetrachloroethene	0.314	0.313	0.3	94	0.00
62 M	Toluene	1.643	1.697	-3.3	95	0.00
63 S	Toluene-d8	1.432	1.433	-0.1	93	0.00
64 M	Chlorobenzene	1.010	1.014	-0.4	94	0.00
65	1,1,1,2-Tetrachloroethane	0.354	0.345	2.5	95	0.00
66 M	Ethyl Benzene	1.792	1.873	-4.5	94	0.00
67 M	m/p-Xylenes	0.688	0.686	0.3	93	0.00
68 M	o-Xylene	0.669	0.647	3.3	91	0.00
69 M	Styrene	1.118	1.098	1.8	93	0.00
70	Isopropylbenzene	1.733	1.805	-4.2	93	0.00
71 M	1,1,2,2-Tetrachloroethane	0.551	0.533	3.3	92	0.00
72	1,2,3-Trichloropropane	0.474	0.500	-5.5	98	0.00
73	Bromobenzene	0.420	0.414	1.4	93	0.00
74	n-propylbenzene	1.998	2.130	-6.6	94	0.00
75	2-Chlorotoluene	1.250	1.236	1.1	93	0.00
76	1,3,5-Trimethylbenzene	1.492	1.521	-1.9	93	0.00
77	t-1,4-Dichloro-2-butene	0.155	0.129	16.8	91	0.00
78	4-Chlorotoluene	1.281	1.268	1.0	93	0.00
79	tert-butylbenzene	1.292	1.279	1.0	94	0.00
80	1,2,4-Trimethylbenzene	1.490	1.538	-3.2	93	0.00
81	sec-Butylbenzene	1.695	1.762	-4.0	93	0.00
82	p-Isopropyltoluene	1.526	1.558	-2.1	93	0.00
83 M	1,3-Dichlorobenzene	0.775	0.750	3.2	93	0.00
84 M	1,4-Dichlorobenzene	0.771	0.732	5.1	92	0.00
85	n-Butylbenzene	1.418	1.447	-2.0	96	0.00
86 T	Hexachloroethane	0.247	0.233	5.7	96	0.00
87 M	1,2-Dichlorobenzene	0.756	0.738	2.4	93	0.00
88	1,2-Dibromo-3-Chloropropane	0.108	0.094	13.0	90	0.00
89	1,2,4-Trichlorobenzene	0.494	0.438	11.3	91	0.00
90	Hexachlorobutadiene	0.239	0.232	2.9	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092619\
Data File : VN058370.D
Acq On : 26 Sep 2019 9:06
Operator : JC/SP
Sample : VSTDCCC020
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC020

Quant Time: Sep 27 03:33:27 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N092419W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Sep 25 01:36:25 2019
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	1.357	1.248	8.0	94	0.00
92	1,2,3-Trichlorobenzene	0.475	0.442	6.9	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0