

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092623\
 Data File : VN079346.D
 Acq On : 26 Sep 2023 10:31
 Operator : JC\MD
 Sample : VN0926WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0926WBL01

Quant Time: Sep 27 06:23:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	154173	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	283276	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	246237	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	71953	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	130060	62.493	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 124.980%	
35) Dibromofluoromethane	8.171	113	103813	55.094	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.180%	
50) Toluene-d8	10.571	98	341663	49.381	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.760%	
62) 4-Bromofluorobenzene	12.853	95	105162	47.650	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.300%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092623\
Data File : VN079346.D
Acq On : 26 Sep 2023 10:31
Operator : JC\MD
Sample : VN0926WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0926WBL01

Quant Time: Sep 27 06:23:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

