

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092623\
 Data File : VN079351.D
 Acq On : 26 Sep 2023 13:44
 Operator : JC\MD
 Sample : 04601-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Sep 27 06:26:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	152390	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	285817	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	259889	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	72671	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	117032	56.891	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.780%
35) Dibromofluoromethane	8.171	113	92990	48.911	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.820%
50) Toluene-d8	10.571	98	328746	47.092	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.180%
62) 4-Bromofluorobenzene	12.853	95	100460	45.115	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.220%
Target Compounds						
						Qvalue
16) Acetone	4.447	43	20043	22.941	ug/l	95
20) Methylene Chloride	5.294	84	9563	4.498	ug/l #	89
43) Isopropyl Acetate	8.694	43	128782	27.002	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092623\
Data File : VN079351.D
Acq On : 26 Sep 2023 13:44
Operator : JC\MD
Sample : 04601-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Sep 27 06:26:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

