

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092624\
 Data File : VN084177.D
 Acq On : 26 Sep 2024 15:18
 Operator : JC\MD
 Sample : VN0926WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0926WBL01

Quant Time: Sep 27 01:10:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 06 23:42:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	28599	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	150635	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	138663	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	75236	28.780	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	95.933%
60) 4-Bromofluorobenzene	12.847	95	59055	24.711	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	82.367%
63) Toluene-d8	10.565	98	188509	29.360	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	97.867%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092624\
Data File : VN084177.D
Acq On : 26 Sep 2024 15:18
Operator : JC\MD
Sample : VN0926WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0926WBL01

Quant Time: Sep 27 01:10:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Sep 06 23:42:47 2024
Response via : Initial Calibration

