

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092625\
 Data File : VN087963.D
 Acq On : 26 Sep 2025 13:49
 Operator : JC\MD
 Sample : VN0926WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0926WBL01

Quant Time: Sep 27 01:24:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	239293	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	535067	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	507778	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	233948	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	235234	58.412	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.820%
35) Dibromofluoromethane	8.153	113	195893	50.424	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.840%
50) Toluene-d8	10.547	98	687185	50.300	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.600%
62) 4-Bromofluorobenzene	12.829	95	217679	44.576	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092625\
Data File : VN087963.D
Acq On : 26 Sep 2025 13:49
Operator : JC\MD
Sample : VN0926WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0926WBL01

Quant Time: Sep 27 01:24:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration

