

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092625\
 Data File : VN087967.D
 Acq On : 26 Sep 2025 15:27
 Operator : JC\MD
 Sample : Q3181-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-A3-01-G

Quant Time: Sep 27 01:26:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	178955	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	392164	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	414447	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	187569	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	179043	59.450	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 118.900%		
35) Dibromofluoromethane	8.153	113	138134	48.514	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.020%		
50) Toluene-d8	10.547	98	504907	50.425	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.860%		
62) 4-Bromofluorobenzene	12.829	95	172948	48.322	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 96.640%		
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	83772	67.605	ug/l	98
20) Methylene Chloride	5.259	84	17906	7.139	ug/l #	88
25) 2-Butanone	7.471	43	12149	6.415	ug/l #	82
40) Benzene	8.583	78	17022	1.502	ug/l #	90
43) Isopropyl Acetate	8.677	43	11126	1.572	ug/l #	89
95) Naphthalene	15.612	128	1479024	134.130	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092625\
Data File : VN087967.D
Acq On : 26 Sep 2025 15:27
Operator : JC\MD
Sample : Q3181-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-A3-01-G

Quant Time: Sep 27 01:26:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration

