

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092625\
 Data File : VN087970.D
 Acq On : 26 Sep 2025 16:31
 Operator : JC\MD
 Sample : Q3183-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Sep 27 01:28:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	186038	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	417660	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	434715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	197936	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	177640	56.738	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.480%
35) Dibromofluoromethane	8.147	113	150016	49.470	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.940%
50) Toluene-d8	10.547	98	537207	50.376	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.760%
62) 4-Bromofluorobenzene	12.829	95	181935	47.730	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.460%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	30966	24.039	ug/l	90
18) Methyl Acetate	5.018	43	20262	6.285	ug/l	97
20) Methylene Chloride	5.271	84	15845	6.077	ug/l #	82
37) Ethyl Acetate	7.547	43	12913	2.697	ug/l #	96
43) Isopropyl Acetate	8.671	43	15785	2.094	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092625\
Data File : VN087970.D
Acq On : 26 Sep 2025 16:31
Operator : JC\MD
Sample : Q3183-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Sep 27 01:28:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration

