

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092625\
 Data File : VN087972.D
 Acq On : 26 Sep 2025 17:13
 Operator : JC\MD
 Sample : Q3183-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-3

Quant Time: Sep 27 01:29:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	179767	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	399691	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	422262	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	190255	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	174092	57.545	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 115.080%	
35) Dibromofluoromethane	8.153	113	147391	50.790	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.580%	
50) Toluene-d8	10.547	98	514816	50.447	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.900%	
62) 4-Bromofluorobenzene	12.829	95	173446	47.548	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 95.100%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	28772	23.115	ug/l	95
18) Methyl Acetate	5.024	43	18571	5.961	ug/l	94
20) Methylene Chloride	5.277	84	14462	5.740	ug/l	90
37) Ethyl Acetate	7.547	43	9536	2.081	ug/l #	91
43) Isopropyl Acetate	8.677	43	12298	1.705	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092625\
Data File : VN087972.D
Acq On : 26 Sep 2025 17:13
Operator : JC\MD
Sample : Q3183-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-3

Quant Time: Sep 27 01:29:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration

