

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092721\
 Data File : VN068704.D
 Acq On : 27 Sep 2021 14:11
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Sep 28 05:28:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N092721W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Sep 27 14:37:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.662	128	71864	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.971	114	406481	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.746	117	339371	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	183553	27.669	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	92.233%
60) 4-Bromofluorobenzene	12.733	95	136105	22.434	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	74.767%
63) Toluene-d8	10.443	98	494570	31.206	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	104.033%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092721\
Data File : VN068704.D
Acq On : 27 Sep 2021 14:11
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Sep 28 05:28:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N092721W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Sep 27 14:37:57 2021
Response via : Initial Calibration

