

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092722\
 Data File : VN074636.D
 Acq On : 27 Sep 2022 18:53
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Sep 28 08:07:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N092722W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Sep 28 07:48:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.498	128	95476	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.827	114	496927	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.621	117	428373	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.286	65	278938	30.985	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	= 103.267%	
60) 4-Bromofluorobenzene	12.604	95	194012	23.316	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	= 77.733%	
63) Toluene-d8	10.310	98	750754	31.071	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	= 103.567%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092722\
Data File : VN074636.D
Acq On : 27 Sep 2022 18:53
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Sep 28 08:07:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N092722W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Sep 28 07:48:31 2022
Response via : Initial Calibration

