


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092723\
Data File : VN079353.D
Acq On : 27 Sep 2023 08:32
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N092723W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Wed Sep 27 23:50:58 2023



AutoFind: Scans 1852, 1853, 1854; Background Corrected with Scan 1845

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	23.3	5639	PASS
75	95	30	60	55.5	13427	PASS
95	95	100	100	100.0	24205	PASS
96	95	5	9	6.9	1673	PASS
173	174	0.00	2	1.7	320	PASS
174	95	50	100	75.8	18348	PASS
175	174	5	9	8.7	1605	PASS
176	174	95	101	95.6	17539	PASS
177	176	5	9	7.1	1249	PASS