

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092724\
 Data File : VN084198.D
 Acq On : 27 Sep 2024 11:30
 Operator : JC\MD
 Sample : VN0927WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0927WBL01

Quant Time: Sep 28 00:58:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	166313	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	288356	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	245293	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	93786	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	125990	50.240	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.480%		
35) Dibromofluoromethane	8.165	113	97926	53.047	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.100%		
50) Toluene-d8	10.565	98	349731	54.801	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 109.600%		
62) 4-Bromofluorobenzene	12.847	95	110064	41.802	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 83.600%		

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092724\
Data File : VN084198.D
Acq On : 27 Sep 2024 11:30
Operator : JC\MD
Sample : VN0927WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0927WBL01

Quant Time: Sep 28 00:58:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

