

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092724\
 Data File : VN084202.D
 Acq On : 27 Sep 2024 14:42
 Operator : JC\MD
 Sample : PB163689ZHE#1
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB163689ZHE#1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/30/2024
 Supervised By :Mahesh Dadoda 09/30/2024

Quant Time: Sep 28 01:00:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	160346	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	292339	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	259781	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	100979	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	121718	50.343	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.680%			
35) Dibromofluoromethane	8.171	113	94185	50.325	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.660%			
50) Toluene-d8	10.565	98	359243	55.525	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 111.040%			
62) 4-Bromofluorobenzene	12.847	95	116237	43.545	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 87.100%			
Target Compounds						
16) Acetone	4.430	43	57984	59.825	ug/l	93
20) Methylene Chloride	5.271	84	7542	3.803	ug/l	89
25) 2-Butanone	7.494	43	8196	6.434	ug/l	96
43) Isopropyl Acetate	8.688	43	170645m	36.829	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092724\
Data File : VN084202.D
Acq On : 27 Sep 2024 14:42
Operator : JC\MD
Sample : PB163689ZHE#1
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB163689ZHE#1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/30/2024
Supervised By :Mahesh Dadoda 09/30/2024

Quant Time: Sep 28 01:00:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

