

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092724\
 Data File : VN084206.D
 Acq On : 27 Sep 2024 16:18
 Operator : JC\MD
 Sample : PB163689ZHE#3
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB163689ZHE#3

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/30/2024
 Supervised By :Mahesh Dadoda 09/30/2024

Quant Time: Sep 28 01:00:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	180778	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	320374	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	275905	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	116965	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	131695	48.313	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.620%		
35) Dibromofluoromethane	8.165	113	100287	48.897	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.800%		
50) Toluene-d8	10.565	98	383010	54.018	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	108.040%		
62) 4-Bromofluorobenzene	12.847	95	129421	44.242	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	88.480%		
Target Compounds						
16) Acetone	4.430	43	63562	58.168	ug/l	100
20) Methylene Chloride	5.283	84	9117	4.077	ug/l	96
25) 2-Butanone	7.494	43	9036	6.291	ug/l	96
43) Isopropyl Acetate	8.688	43	230746m	45.442	ug/l	
51) 4-Methyl-2-Pentanone	10.441	43	14737m	5.332	ug/l	
95) Naphthalene	15.641	128	59172	9.136	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092724\
Data File : VN084206.D
Acq On : 27 Sep 2024 16:18
Operator : JC\MD
Sample : PB163689ZHE#3
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB163689ZHE#3

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/30/2024
Supervised By :Mahesh Dadoda 09/30/2024

Quant Time: Sep 28 01:00:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

